


Issued by NMI Certin B.V.,  
designated and notified by the Netherlands to perform tasks with respect to conformity assessment procedures mentioned in article 17 of Directive 2014/32/EU, after having established that the measuring instrument meets the applicable requirements of Directive 2014/32/EU, to:

Manufacturer Smartmeters Technologies Private Limited  
Plot No. 268, Opp Prince Spintex Pvt. Ltd.  
Bagodara – Dholka Road,  
Gangad, Bavla, Ahmedabad-382240  
Gujarat, India

Measuring instrument **Diaphragm gas meter**  
Type : ALPS G1.6  
EVEREST G1.6

Manufacturer's mark or name :   
*Intelligence inside, Smart outside*

Destined for the measurement of : Gas volume  
Accuracy class : Class 1,5  
Environment classes : M1 / E1  
Ambient temperature range : -10 °C / +55 °C  
Gas temperature range : -10 °C / +40 °C  
Designed for : Condensing humidity  
Location : Closed

Further properties are described in the annexes:  
– Description T12332 revision 1;  
– Documentation folder T12332-2.

Valid until 5 September 2032

Initially issued 5 September 2022

Remark This revision replaces the earlier version(s), including its documentation folder.

Issuing Authority **NMI Certin B.V., Notified Body number 0122**  
18 July 2024

## Certification Board

**NMI Certin B.V.**  
Thijsseweg 11  
2629 JA Delft  
The Netherlands  
T +31 88 636 2332  
[certin@nmi.nl](mailto:certin@nmi.nl)  
[www.nmi.nl](http://www.nmi.nl)

This document is issued under the provision that no liability is accepted and that the manufacturer shall indemnify third-party liability.

The designation of NMI Certin B.V. as Notified Body can be verified at <http://ec.europa.eu/growth/tools-databases/nando/>

Reproduction of the complete document only is permitted.

This document is digitally signed and sealed. The digital signature can be verified in the blue ribbon at the top of the electronic version of this certificate.