

Issued by NMI Certin B.V.,
designated and notified by the Netherlands to perform tasks with respect to conformity assessment procedures mentioned in article 17 of Directive 2014/32/EU, after having established that the measuring instrument meets the applicable requirements of Directive 2014/32/EU, to:

Manufacturer Honeywell Automation India Ltd.
Plot No. 3, Gat No. 181
Village Fulgaon, Tal-Haveli
Pune 412216, Maharashtra
India

Measuring instrument An **electronic gas-volume conversion device (EVCD)**, intended to be used for gas volume conversion as a sub-assembly (according to article 4 of the MID) of a gas meter.

Type : EK205
EK205-M

Manufacturer's mark or name : Elster

Conversion principle : T, PT or PTZ

Ambient temperature range : -25 °C / +55 °C

Designed for : non-condensing humidity

Environment classes : M2 / E2

The intended location for the instrument is open.

Further properties are described in the annexes:

- Description T12625 revision 1;
- Documentation folder T12625-2.

Valid until 18 July 2033

Initially issued 18 July 2023

Remark This revision replaces the earlier version(s), including its documentation folder.

Issuing Authority **NMI Certin B.V., Notified Body number 0122**
8 May 2025

Certification Board