

Issued by NMI Certin B.V.,
designated and notified by the Netherlands to perform tasks with respect to conformity assessment procedures mentioned in article 17 of Directive 2014/32/EU, after having established that the measuring instrument meets the applicable requirements of Directive 2014/32/EU, to:

Manufacturer CPL Concordia Soc. Coop.
Via A. Grandi, 39
41033 Concordia s/S (MO)
Italia

Measuring instrument An **electronic gas-volume conversion device (EVCD)**, intended to be used for gas volume conversion as a sub-assembly (according to article 4 of the MID) of a gas meter.

Type : FlyCOR
Manufacturer's mark or name : CPL Concordia
Conversion principle : T, PT or PTZ
Ambient temperature range : -25 °C / +70 °C
Designed for : condensing humidity
Environment classes : M2 / E2
The intended location for the instrument is open.

Further properties are described in the annexes:

- Description T12820 revision 0;
- Documentation folder T12820-1.

Valid until 23 May 2034

Initially issued 23 May 2024

Issuing Authority

NMI Certin B.V., Notified Body number 0122
28 May 2024

Certification Board

NMI Certin B.V.
Thijsseseweg 11
2629 JA Delft
The Netherlands
T +31 88 636 2332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the manufacturer shall indemnify third-party liability.

The designation of NMI Certin B.V. as Notified Body can be verified at <http://ec.europa.eu/growth/tools-databases/nando/>

Reproduction of the complete document only is permitted.

This document is digitally signed and sealed. The digital signature can be verified in the blue ribbon at the top of the electronic version of this certificate.