

Issued by NMI Certin B.V.,
designated and notified by the Netherlands to perform tasks with respect to conformity assessment procedures mentioned in article 17 of Directive 2014/32/EU, after having established that the measuring instrument meets the applicable requirements of Directive 2014/32/EU, to:

Manufacturer FlowM Technologies OU
Narva mnt 5, 10117 Tallinn,
Harju maakond, Estonia

Measuring instrument **A Rotary Displacement Gas Meter**

Type	: RM Series
Manufacturer's mark or name	: FlowM
Destined for the measurement of	: Gas volume
Accuracy class	: Class 1,0
Environment classes	: M1
Temperature range	: Depending on Q_{min} , see table 1.2.1
Maximum pressure	: 20 bar(g)

Further properties are described in the annexes:

- Description T12989 revision 0;
- Documentation folder T12989-1.

Valid until 18 February 2035

Initially issued 18 February 2025

Issuing Authority **NMI Certin B.V., Notified Body number 0122**
18 February 2025

Certification Board