

CERTIFICATE

EC-Type examination certificate 08-4041 revision A

Authorised representative : GE Power Controls
Contact person : Sergio Bueno
Address : AV Camara de la Industria 9 - PG Industrial N° 1
Postal code, Place : 28938, Mostoles - Madrid
Country : Spain

Manufacturer : EMCO Limited
Address : Plot no. F-5, Road no. 28, Wagle Industrial Estate
Postal code, Place : 400604, Thane
Country : India

Instrument : Electronic single-phase two-wire energy meter
Direct connected

Type : RS 1978
Register : LCD
Accuracy Class : 1 / B
Measurement range : 230 V
10(60)A
50 Hz
3200 imp./kWh

Temperature range : -25..55 °C
Use : Indoor
Protection Class : II

The energy meter meets the requirements of Directive 2004/22/EC of the European parliament and the council of 31 March 2004 on measuring instruments.


Certification was based on compliance with the following harmonised standards:

EN 50470-1 (2006) : Electricity metering equipment (a.c.)-part 1: General requirements, tests and test conditions - Metering equipment (class indexes A, B and C)
EN 50470-3 (2006) : Electricity metering equipment (a.c.)-part 3: Particular requirements - Static meters for active energy (class indexes A, B and C)

Valid until : December 8, 2018

The results are recorded in the following annex: test report 2114361-QUA/CAL 08-4041A.

KEMA Quality B.V.
Arnhem, December 8, 2008



ing. J.H.M. Overbeek
Certification manager
Notified body number 0344



ir. P.J.J.G. Nabuurs
Managing Director

Integral publication of this certificate and adjoining reports is allowed.

KEMA Quality B.V. Utrechtseweg 310, 6812 AR Arnhem P.O. Box 5185, 6802 ED Arnhem The Netherlands
T +31 26 3 56 20 00 F +31 26 3 52 58 00 customer@kema.com www.kema.com Registered Arnhem 09085396