

EU-TYPE EXAMINATION CERTIFICATE

Certificate number	1010-21
Manufacturer	CLOU Global Technology Co., Ltd.,
Contact person	Ms. Alice Liu
Address	14th Floor, Block A, CLOU Building, Baoshen Road, Songpingshan Community, Xili Street, Nanshan District, 518057 Shenzhen
Country	China
Production site	CLOU Global Technology Co., Ltd.,
Address	Clou Industrial Park, Baolong road, Longgang district, Shenzhen,
Country	China
Instrument	Electronic single-phase two-wire, two-element energy meter Direct connected
Mark - Type	CL710K22/CL710K24
Register	LCD
Accuracy Class	Active: B;
Measurement range	120 – 240 V, 0,25...5(100) A 50 and 60 Hz 1000 imp./kWh
Temperature range	-40 .. 70 °C
Use	Indoor
Protection Class	II
Environmental class	M1, M2, E1 and E2
Registry method	Bi-directional method separate registers: received- and delivered energy of the whole connection is added in separate registers

The energy meters meet the requirements of Directive 2014/32/EU, on the harmonisation of the laws of Member States relating to the making available on the market of measuring instruments (recast).


Certification was based on compliance with the following harmonised standards:

EN 50470-1:2006	Electricity metering equipment (a.c.)-part 1: General requirements, tests and test conditions - Metering equipment (class indexes A, B and C)
EN 50470-3:2006	Electricity metering equipment (a.c.)-part 3: Particular requirements - Static meters for active energy (class indexes A, B and C)
Valid until	19 January 2032

Based on a non-recurrent examination.

Revision 1 of this certificate. The results are recorded in the following annex: test report 1508-22 R0.

KEMA B.V.



Bas Verhoeven
Director, High-Voltage
Laboratory

Arnhem, 19 January 2022