



CERTIFICATE

EC-Type examination certificate TIC 1070-14

Manufacturer	: Inepro Metering B.V.
Manufacturer's name	: Inepro Metering, Inepro, ENX
Contact person	: D. van der Vaart
Address	: Pondweg 7
Postal code, Place	: 2153PK, Nieuw-Vennep
Country	: The Netherlands
Production site	: Zhejiang Yongtailong Electronic Co., Ltd.
Address	: No. 320 Tongsheng Road, Tongxiang
Postal code, Place	: 314500, Jiaxing City, Zhejiang
Country	: China
Instrument	: Electronic single-phase two-wire energy meter Direct connected
Mark - Type	: MEMO4-M, MEMO4-M 2T, MEMO4-M Mb and MEMO4-M Mod
Register	: LCD
Accuracy Class	: 1 / B
Measurement range	: 230 V 5(45) A 50 Hz 10000 imp./kWh
Temperature range	: -25 .. 55 °C
Use	: Indoor
Protection Class	: II
Environmental class	: M1, E1
Registry method	: Programmable

The energy meter meets the requirements of Directive 2004/22/EC of the European parliament and the council of 31 March 2004 on measuring instruments.

Certification was based on compliance with the following harmonised standards:


EN 50470-1 (2006)	: Electricity metering equipment (a.c.)-part 1: General requirements, tests and test conditions - Metering equipment (class indexes A, B and C)
EN 50470-3 (2006)	: Electricity metering equipment (a.c.)-part 3: Particular requirements - Static meters for active energy (class indexes A, B and C)

Valid until : 17 February 2025

Based on a non-recurrent examination.

Version 3 of this certificate. The results are recorded in the following annex: test report TIC 1070-14.

KEMA Nederland B.V.
Arnhem, 12 March 2015


ing. S.A.M. Verhoeven
Business Line Director Testing, Inspections & Certification
Notified body number 2290