

Issued by NMI Certin B.V.
Hugo de Grootplein 1
3314 EG Dordrecht
The Netherlands

Notified Body number 0122

In accordance with The "Metrologiewet" (Stb. 2006, 137) as Dutch implementation of Directive 2004/22/EC on measuring instruments (MID).

Applicant Landis+Gyr AG
Feldstrasse 1
CH 6301 Zug
Switzerland

In respect of A model of a static **Active Electrical Energy Meter**
Type : Z.D100AS

Characteristics Reference voltage : 3x230/400 V or 3x230 V
Reference current : 5, 10 or 20 A
Destined for the measurement of : electrical energy, in a
- three-phase four-wire network (ZMD)
- three-phase three-wire network (ZFD)
- two-phase three-wire network (ZMD)
- single-phase two-wire network (ZMD)
Accuracy class : A (Z.D) or B (ZMD)
Mechanical environment class : M1
Electromagnetic environment class : E2
Temperature range : -40°C / +70°C

In the description number T10027 revision 2 further characteristics are described.

Valid until 12 March 2017

Description and documentation The instrument is described in the description number T10027 revision 2 and documented in the documentation folder T10027-1, appertaining to this EC-type examination certificate.

Remark This version replaces the earlier version, except for its documentation folder.

Dordrecht, 8 March 2011
NMI Certin B.V.


P.A. van Breugel
Director