


Issued by	NMi Certin B.V., designated and notified by the Netherlands to perform tasks with respect to conformity modules mentioned in article 17 of Directive 2014/32/EU, after having established that the Measuring instrument meets the applicable requirements of Directive 2014/32/EU, to:
Manufacturer	Itron (Actaris) 1 Avenue des Temps Modernes ZI du Bernais B.P. 23 86361 Chasseneuil du Poitou France
Measuring instrument	A static Active Electrical Energy Meter Type : ACE661B / ACE661D and ACE662B / ACE662D Reference voltage : 3x57,7/100 V 3x240/415 V Reference current : ACE66.B: 1; 1,5; 2,5 or 5 A ACE66.D: 5; 10; 15 or 20 A Destined for the measurement of : electrical energy, in a - three-phase four-wire network - three-phase three-wire network Accuracy class : ACE66.B: B or C ACE66.D: A or B Environment classes : M2 / E2 Temperature range : -40 °C / +70 °C Further properties are described in the annexes: - Description T10053 revision 25; - Documentation folder T10053-12.
Valid until	28 March 2017
Remarks	This revision replaces the earlier versions, except for its documentation folder.

Issuing Authority **NMi Certin B.V., Notified Body number 0122**
9 June 2016

C. Oosterman
Head Certification Board