

Nederlands Meetinstituut

EC-type examination certificate

Number T10062 revision 2 Project number 609879 Page 1 of 1

Issued by

NMi Certin B.V.

Hugo de Grootplein 1 3314 EG Dordrecht The Netherlands

Notified Body number 0122

In accordance with

The "Metrologiewet" (Stb. 2006, 137) as Dutch implementation of Directive

2004/22/EC on measuring instruments (MID).

Manufacturer

ELSTER Messtechnik GmbH

Otto-Hahn-Strasse 25 68263 Lampertheim

Germany

In respect of

A model of a static Active Electrical Energy Meter

Type

: A1350

Characteristics

Reference voltage

: 3 x 230/400 V (direct & CT connected)

3 x 57,7/100 V (CT connected) : 5 or 10 A (direct connected)

Reference current

1 or 5 A (CT connected)

Destined for the measurement of

: electrical energy, in a

- three-phase four-wire network - three-phase three-wire network

Accuracy class

: B or A (direct connected) B (CT connected, reference current 1 A)

C or B (CT connected, reference current 5 A)

Environment classes Temperature range

: M1 / E2 : -40°C / +70°C

In the description number T10062 revision 2 further characteristics are described.

Valid until

8 May 2017

Description and The instrument is described in the description number T10062 revision 2 documentation and documented in the documentation folder T10062-1, appertaining to this

EC-type examination certificate.

Remark

This revision replaces the previous version, except for its documentation folder.

Dordrecht, 31 October 2008 NMi Certin B.

Ing. C. Oosterman

Manager Product Certification

NMi Certin BV

Hugo de Grootplein 1 3314 EG Dordrecht PO Box 394 3300 AJ Dordrecht, NL T +31 78 6332332 F+31 78 6332309 certin@nmi.nl www.nmi.nl

Parties concerned can lodge objection against this decision, within six weeks after the date of submission, to the general manager of NMi (see "Regulation objection and appeal against decisions of NMi")

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

Reproduction of the complete document is permitted.

