



EC-type examination certificate

Number **T10065** revision 0
Project number 700386
Page 1 of 1

Issued by NMI Certin B.V.
Hugo de Grootplein 1
3314 EG Dordrecht
The Netherlands

Notified Body number 0122

In accordance with The "Metrologiewet" (Stb. 2006, 137) as Dutch implementation of Directive 2004/22/EC on measuring instruments (MID).

Applicant Landis+Gyr AG
Feldstrasse 1
CH 6301 Zug
Switzerland

In respect of A model of a static **Active Electrical Energy Meter**

Type : ZCF100Ax

Characteristics	Reference voltage	: 230 V
	Reference current	: 5 or 10 A
	Destined for the measurement of	: electrical energy, in a - one-phase two-wire network
	Accuracy class	: A or B
	Mechanical environment class	: M1
	Electromagnetic environment class	: E2
	Temperature range	: -40°C / +70°C

In the description number T10065 revision 0 further characteristics are described.

Valid until 4 April 2017

Description and documentation The instrument is described in the description number T10065 revision 0 and documented in the documentation folder T10065-1, appertaining to this EC-type examination certificate.

Dordrecht, 4 April 2007
NMI Certin B.V.

Ing. C. Oosterman
Manager Product Certification

