

Issued by NMI Certin B.V.,
designated and notified by the Netherlands to perform tasks with respect to
conformity modules mentioned in article 9 of Directive 2004/22/EC, after
having established that the Measuring instrument meets the applicable
requirements of Directive 2004/22/EC, to:

Manufacturer Itron
1 Avenue des temps modernes
ZI du Bernais B.P. 23
86361 Chasseneuil du Poitou
France

Measuring instrument **A static Active Electrical Energy Meter**
Type : SL761B / SL761E / SL760
Reference voltage : 3x57,7/100 3x240/415 V
Reference current : SL761B or SL760 transformer connected (TC):
1; 1,5; 2,5 or 5A
SL761E or SL760 direct connected (DC):
5; 10; 15 or 20A
Destined for the measurement of : electrical energy, in a
- three-phase four-wire network
- three-phase three-wire network
Accuracy class : SL761B or SL760 (TC): B or C
SL761E or SL760 (DC): A or B
Environment classes : M2 / E2
Temperature range : -40 °C / +70 °C

Further properties are described in the annexes:
– Description T10090 revision 10;
– Documentation folder T10090-3.

Valid until 24 September 2017

Remarks This revision replaces the earlier versions, except for its documentation
folder.

Issuing Authority **NMI Certin B.V., Notified Body number 0122**
11 June 2013

C. Oosterman
Head Certification Board

NMI Certin B.V.
Hugo de Grootplein 1
3314 EG Dordrecht
The Netherlands
T +31 78 6332332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision
that no liability is accepted and that the
applicant shall indemnify third-party liability.

The designation of NMI Certin BV.as Notified
Body can be verified at [http://
ec.europa.eu/enterprise/newapproach/hando/](http://ec.europa.eu/enterprise/newapproach/hando/)

Parties concerned can lodge
objection against this decision,
within six weeks after the date of
submission, to the general manager
of NMI (see www.nmi.nl).

Reproduction of the complete
document only is permitted.