

## EC type-examination certificate

Number **T10111** revision 3 Project number 10200380 Page 1 of 1

Issued by NMi Certin B.V.,

designated and notified by the Netherlands to perform tasks with respect to conformity modules mentioned in article 9 of Directive 2004/22/EC, after having established that the Measuring instrument meets the applicable

requirements of Directive 2004/22/EC, to:

Manufacturer ENEL Distribuzione SpA

00198 ROMA Via Ombrone 2

Italy

Measuring instrument A static Active Electrical Energy Meter

Type + + + + + + + + + + : + GEM

Reference voltage : 230 V Reference current : 10 A

Destined for the measurement of : electrical energy, in a

- single-phase two-wire network

Accuracy class : B

Environment classes : M1 / E2

Temperature range : -40 °C / +70 °C

Further properties are described in the annexes

Description T10111 revision 3Documentation folder T10111-2

Valid until 21 December 2017

Remarks This revision replaces the earlier versions, except for its documentation

folder.

Issuing Authority

NMi Certin B.V., Notified Body number 0122

12 October 2010

Costerman

**Head Certification Board** 

NMi Certin B.V. Hugo de Grootplein 1 3314 EG Dordrecht The Netherlands T +31 78 6332332 certin@nmi.nl www.nmi.nl This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The designation of NMi Certin B.V. as Notified Body can be verified at http://ec.europa.eu/enterprise/newapproach/nando/

Parties concerned can lodge objection against this decision, within six weeks after the date of submission, to the general manager of NMi (see www.nmi.nl).

Reproduction of the complete document only is permitted.

