

Issued by	NMi Certin B.V., designated and notified by the Netherlands to perform tasks with respect to conformity modules mentioned in article 9 of Directive 2004/22/EC, after having established that the Measuring instrument meets the applicable requirements of Directive 2004/22/EC, to:
Manufacturer	Iskraemeco Savska loka 4 4000 Kranj Slovenia
Measuring instrument	<b>A static Active Electrical Energy Meter</b> Type : MT371, MT372 or MT375 Reference voltage : 3 x 230/400 V Reference current : 5 or 10 A (direct connected) 5 A (indirect connected) Destined for the measurement of : electrical energy, in a - three-phase four-wire network  Accuracy class : A or B Environment classes : M1 / E2 Temperature range : -40 °C / +70 °C  Further properties are described in the annexes: - Description T10118 revision 5; - Documentation folder T10118-4.
Valid until	5 March 2018
Remarks	This revision replaces the earlier versions, including its documentation folder.

Issuing Authority **NMi Certin B.V., Notified Body number 0122**  
26 March 2015

  
C. Oosterman  
Head Certification Board