


Issued by	NMi Certin B.V., designated and notified by the Netherlands to perform tasks with respect to conformity modules mentioned in article 9 of Directive 2004/22/EC, after having established that the Measuring instrument meets the applicable requirements of Directive 2004/22/EC, to:
Manufacturer	EMH metering GmbH & Co. KG Südring 5 D-19243 Wittenburg Germany
Measuring instrument	A static Active Electrical Energy Meter Type : ED1xxx / ED2xxx Reference voltage : 220, 230 or 240 V Reference current : 5 or 10 A Destined for the measurement of : electrical energy, in a - single-phase two-wire network Accuracy class : A or B Environment classes : M1 / E2 Temperature range : -40 °C / +70 °C Further properties are described in the annexes – Description T10177 revision 6 – Documentation folder T10177-4
Valid until	20 January 2019
Remarks	This revision replaces the earlier versions, except for its documentation folder.

Issuing Authority	NMi Certin B.V., Notified Body number 0122 23 May 2011  C. Oosterman Head Certification Board
-------------------	---