



EC type-examination Certificate

Number **T10181** revision 1
Project number SO12200788
Page 1 of 1

Issued by	NMi Certin B.V., designated and notified by the Netherlands to perform tasks with respect to conformity modules mentioned in article 9 of Directive 2004/22/EC, after having established that the Measuring instrument meets the applicable requirements of Directive 2004/22/EC, to:
Manufacturer	ACEAN Compteurs ZI de la Liane Sud – BP439 62206 Boulogne sur mer France
Measuring instrument	A static Active Electrical Energy Meter Type : WH6165 Reference voltage : 220, 230 or 240 V Reference current : 10 A Destined for the measurement of : electrical energy, in a single-phase two-wire network Accuracy class : A or B Mechanical environment class : M1 Electromagnetic environment class : E2 Temperature range : -40 °C / +70 °C Further properties are described in the annexes – Description T10181 revision 1 – Documentation folder T10181-1
Valid until	19 February 2019
Remark	This revision replaces the earlier version, except for its documentation folder.
Issuing Authority	NMi Certin B.V., Notified Body number 0122 19 October 2012 C. Oosterman Head Certification Board

NMi Certin B.V.
Hugo de Grootplein 1
3314 EG Dordrecht
The Netherlands
T +31 78 6332332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The designation of NMi Certin B.V. as Notified Body can be verified at <http://ec.europa.eu/enterprise/newapproach/nando/>

Parties concerned can lodge objection against this decision, within six weeks after the date of submission, to the general manager of NMi (see www.nmi.nl).

Reproduction of the complete document only is permitted.

