

Issued by NMI Certin B.V.,  
 designated and notified by the Netherlands to perform tasks with respect to  
 conformity modules mentioned in article 17 of Directive 2014/32/EU, after  
 having established that the Measuring instrument meets the applicable  
 requirements of Directive 2014/32/EU, to:

Manufacturer Elster GmbH  
 Otto-Hahn-Strasse 25  
 68623 Lampertheim  
 Germany

Measuring instrument **A static Active Electrical Energy Meter**

Type : AS1253-D1... (DC) or  
 AS1253-W1... (CT)

Manufacturer's mark or name : Elster or Honeywell

Reference voltage : 3x230/400V

Reference current : 5, 10 or 20 A (DC)  
 1, 2 or 5 A (CT)

Destined for the measurement of : electrical energy, in a  
 - three-phase four-wire network  
 - single-phase two-wire network

Accuracy class : A or B (DC)  
 B or C (CT)

Environment classes : M1 / E2

Temperature range : -40 °C / +70 °C

Further properties are described in the annexes  
 - Description T10190 revision 3  
 - Documentation folder T10190-3

Valid until 30 March 2019

Remarks This revision replaces the earlier versions, including its documentation folder.

Issuing Authority **NMI Certin B.V., Notified Body number 0122**  
 1 March 2018

  
 C. Oosterman  
 Head Certification Board

**NMI Certin B.V.**  
 Hugo de Grootplein 1  
 3314 EG Dordrecht  
 The Netherlands  
 T +31 78 6332332  
 certin@nmi.nl  
 www.nmi.nl

This document is issued under the provision  
 that no liability is accepted and that the  
 applicant shall indemnify third-party liability.

The designation of NMI Certin BV.as Notified  
 Body can be verified at [http://  
 ec.europa.eu/growth/tools-databases/nando/](http://ec.europa.eu/growth/tools-databases/nando/)

Reproduction of the complete  
 document only is permitted.