

EC type-examination Certificate

Number **T10202** revision 3 Project number 10200915 Page 1 of 1

Issued by NMi Certin B.V.,

designated and notified by the Netherlands to perform tasks with respect to conformity modules mentioned in article 9 of Directive 2004/22/EC, after having established that the Measuring instrument meets the applicable

requirements of module B of Directive 2004/22/EC, to:

Manufacturer ABB AB / Cewe-Control

Arnöleden 2

SE-61129 Nyköping

Sweden

Measuring instrument A static Active Electrical Energy Meter

Type : DELTAmax Type: D.... Trademark: ABB

Company: ABB, Sweden

DELTAmax Type: E Trademark: ENDYS Company: Hochhuth Gmbh, Germany DELTAmax Type: B Trademark: BERG

Company: Berg Energiekontrolle Gmbh Germany DELTAmax Type: R Trademark: Rauscher & Stoecklin

Company: Rauscher & Stoecklin AG Switzerland

Reference voltage 🗼 💠 : 3 x 57,7/100 V ... 3 x 288/500 V

Reference current : 5 A (direct connected)

1 A (indirect connected)

Destined for the : electrical energy, in a

measurement of - three-phase four-wire network

three-phase three-wire network
single-phase two-wire network

Accuracy class : A or B
Environment classes : M2 / E2

Temperature range : -40 °C / +70 °C

Further properties are described in the annexes

Description T10202 revision 3
Documentation folder T10202-3

Valid until 6 July 2019

Remarks This revision replaces the earlier versions including its documentation folder

Issuing Authority NMi Certin B.V., Notified Body number 0122

30 November 2010

C. Oosterman

Head Certification Board

NMi Certin B.V. Hugo de Grootplein 1 3314 EG Dordrecht The Netherlands T +31 78 6332332 certin@nmi.nl This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The designation of NMi Certin BV.as Notified Body can be verified at http://ec.europa.eu/enterprise/newapproach/nando/

Parties concerned can lodge objection against this decision, within six weeks after the date of submission, to the general manager of NMi (see www.nmi.nl).

Reproduction of the complete document only is permitted.

