

Issued by	NMi Certin B.V., designated and notified by the Netherlands to perform tasks with respect to conformity modules mentioned in article 9 of Directive 2004/22/EC, after having established that the Measuring instrument meets the applicable requirements of Directive 2004/22/EC, to:
Manufacturer	Kamstrup A/S Industrivej 28, Stilling DK-8660 Skanderborg Denmark
Measuring instrument	<p><b>A static <b>Active Electrical Energy Meter</b></b></p> Type : 685-x5x-Bx-3x-xxx (Kamstrup 351B) Reference voltage : 3x230/400V Reference current : 5 A Destined for the measurement of : electrical energy, in a - three-phase four-wire network - three-phase three-wire network Accuracy class : B or C Environment classes : M1 / E2 Temperature range : -40 °C / +70 °C
	Further properties are described in the annexes - Description T10224 revision 7 - Documentation folder T10224-2
Valid until	18 August 2019
Remarks	This revision replaces the earlier versions, except for its documentation folder.

Issuing Authority **NMi Certin B.V., Notified Body number 0122**  
24 April 2012

C. Oosterman  
Head Certification Board