

Issued by	NMi Certin B.V., designated and notified by the Netherlands to perform tasks with respect to conformity modules mentioned in article 9 of Directive 2004/22/EC, after having established that the Measuring instrument meets the applicable requirements of Directive 2004/22/EC, to:
Manufacturer	Landis+Gyr 30 Avenue du Président Auriol BP 3150 03115 Montluçon Cedex France
Measuring instrument	A static <b>Active Electrical Energy Meter</b> Type : L30C1 Reference voltage : 3x 230 V / 400 V Reference current : 10 A Destined for the measurement of : electrical energy, in a - three-phase four-wire network Accuracy class : B Environment classes : M2 / E2 Temperature range : -40 °C / +70 °C  Further properties are described in the annexes - Description T10267 revision 1 - Documentation folder T10267-2
Valid until	1 April 2020
Remarks	This revision replaces the earlier version, including its documentation folder.

Issuing Authority **NMi Certin B.V., Notified Body number 0122**  
24 February 2011

  
C. Oosterman  
Head Certification Board