

Issued by NMI Certin B.V.,  
designated and notified by the Netherlands to perform tasks with respect to  
conformity modules mentioned in article 17 of Directive 2014/32/EU, after  
having established that the Measuring instrument meets the applicable  
requirements of Directive 2014/32/EU, to:

Manufacturer Networked Energy Services  
5215 Hellyer Avenue, Suite 150  
San Jose, CA 95138  
United States of America

Measuring instrument A static **Active Electrical Energy Meter**

Type	: 83332-1xxxx
Reference voltage	: 230 V
Reference current	: 5 A
Destined for the measurement of	: electrical energy, in a - single-phase two-wire network
Accuracy class	: A or B
Environment classes	: M2 / E2
Temperature range	: -40 °C / +70 °C

Further properties are described in the annexes:

- Description T10322 revision 12;
- Documentation folder T10322-3.

Valid until 2 February 2021

Remarks This revision replaces the earlier versions, except for its documentation folder.

Issuing Authority

**NMI Certin B.V., Notified Body number 0122**  
12 March 2019

  
C. Oosterman  
Head Certification Board