

EC type-examination Certificate

Number **T10336** revision 3
Project number 12200309
Page 1 of 1

| | |
|----------------------|---|
| Issued by | NMi Certin B.V., designated and notified by the Netherlands to perform tasks with respect to conformity modules mentioned in article 9 of Directive 2004/22/EC, after having established that the Measuring instrument meets the applicable requirements of Directive 2004/22/EC, to: |
| Manufacturer | Landis+Gyr 1 Lysander Drive, Northfields Industrial Estate, Market Deeping, Peterborough, PE6 8FB England |
| Measuring instrument | A static Active Electrical Energy Meter Type : (E450) ZMX310G and ZMX320G Reference voltage : 3x230/400 V Reference current : 5 A or 10 A Destined for the measurement of : electrical energy, in a - three-phase four-wire network - two-phase three-wire network - single-phase two-wire network - three-phase three-wire network Accuracy class : A (ZMX320) or B (ZMX310) Environment classes : M2 / E2 Temperature range : -40 °C / +55 °C Further properties are described in the annexes - Description T10336 revision 3 - Documentation folder T10336-4 |
| Valid until | 18 February 2021 |
| Remark | This revision replaces the earlier version, including its documentation folder. |

Issuing Authority **NMi Certin B.V., Notified Body number 0122**
26 June 2012

C. Oosterman
Head Certification Board

NMi Certin B.V.
Hugo de Grootplein 1
3314 EG Dordrecht
The Netherlands
T +31 78 6332332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The designation of NMi Certin B.V. as Notified Body can be verified at <http://ec.europa.eu/enterprise/newapproach/nando/>

Parties concerned can lodge objection against this decision, within six weeks after the date of submission, to the general manager of NMi (see www.nmi.nl).

Reproduction of the complete document only is permitted.