


Issued by	NMi Certin B.V., designated and notified by the Netherlands to perform tasks with respect to conformity modules mentioned in article 9 of Directive 2004/22/EC, after having established that the Measuring instrument meets the applicable requirements of Directive 2004/22/EC, to:
Manufacturer	ELEKTROMED Elektronik San. ve Saglik Hizmetleri A.S. Organize Sanayi Bolgesi Buyuk Selcuklu Bulvari Uygurlar Cad. No:4 06935 - Sincan-Ankara Turkey
Measuring instrument	A static Active Electrical Energy Meter Type : EC058MBW Reference voltage : 230 V Reference current : 5 A Destined for the measurement of : electrical energy, in a - one-phase two-wire network Accuracy class : A or B Environment classes : M2 / E2 Temperature range : -40 °C / +70 °C Further properties are described in the annexes – Description T10356 revision 0 – Documentation folder T10356-1
Valid until	14 June 2021

Issuing Authority	NMi Certin B.V., Notified Body number 0122 14 June 2011  C. Oosterman Head Certification Board
-------------------	--