

Issued by	NMi Certin B.V., designated and notified by the Netherlands to perform tasks with respect to conformity modules mentioned in article 17 of Directive 2014/32/EU, after having established that the Measuring instrument meets the applicable requirements of Directive 2014/32/EU, to:
Manufacturer	KOHLER ELEKTRIK SAYACLARI SAN.TIC.A.S. Akcaburgaz Mah. 3055.Sok. No:3 Kirac-Esenyurt 34522 Istanbul Turkey
Measuring instrument	A static Active Electrical Energy Meter Type : AEL.TF.19 and AEL.TF.21 as transformer connect meter (TCM) AEL.TF.20 and AEL.TF.24 as direct connected meter (DCM) Manufacturer's mark or name : Köhler Reference voltage : 3x57,7/100V or 3x230/400V (TCM) 3x230/400V (DCM) Reference current : 1 A (TCM) 5 A (DCM) Destined for the measurement of : electrical energy, in a - three-phase four-wire network Accuracy class : A or B (DCM) B or C (TCM) Environment classes : M1 / E2 Temperature range : -40 °C / +70 °C Further properties are described in the annexes - Description T10392 revision 2 - Documentation folder T10392-3
Valid until	31 October 2021
Remarks	This revision replaces the earlier versions, including its documentation folder.

Issuing Authority **NMi Certin B.V., Notified Body number 0122**
26 April 2017



C. Oosterman
Head Certification Board

NMi Certin B.V.
Hugo de Grootplein 1
3314 EG Dordrecht
The Netherlands
T +31 78 6332332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The designation of NMi Certin B.V. as Notified Body can be verified at <http://ec.europa.eu/enterprise/newapproach/nando/>

Reproduction of the complete document only is permitted.