

Issued by NMI Certin B.V.,
designated and notified by the Netherlands to perform tasks with respect to conformity modules mentioned in article 17 of Directive 2014/32/EU, after having established that the Measuring instrument meets the applicable requirements of Directive 2014/32/EU, to:

Manufacturer Landis + Gyr
1 Lysander Drive, Northfields Ind Est
Peterborough PE6 8FB
Great Britain

Measuring instrument **A static Active Electrical Energy Meter**

Type : E450 ZCXi1x0
(E450) ZCX110 and ZCX120

Manufacturer's mark or name : Landis + Gyr

Reference voltage : 230 V

Reference current : 5 A or 10 A

Destined for the measurement of : electrical energy, in a
- single-phase two-wire network

Accuracy class : A or B

Environment classes : M2 / E2

Temperature range : -40 °C / +70 °C

Further properties are described in the annexes:
- Description T10527 revision 5;
- Documentation folder T10527-2.

Valid until 30 October 2022

Remarks This revision replaces the earlier versions, except for its documentation folder.

Issuing Authority

NMI Certin B.V., Notified Body number 0122
29 July 2019


C. Oosterman
Head Certification Board