

Issued by NMI Certin B.V.,
designated and notified by the Netherlands to perform tasks with respect to conformity modules mentioned in article 17 of Directive 2014/32/EU, after having established that the Measuring instrument meets the applicable requirements of Directive 2014/32/EU, to:

Manufacturer EDM Europe Ltd
1 Gosforth Park Avenue,
Gosforth Business Park
NE12 8EG Newcastle upon Tyne
United Kingdom

Measuring instrument A static **Active Electrical Energy Meter**

Type	:	ES-10A, Mk7G-4T
Manufacturer's mark or name	:	EDMI
Reference voltage	:	230 V
Reference current	:	5 A
Destined for the measurement of	:	electrical energy, in a - single-phase two-wire network
Accuracy class	:	A or B
Environment classes	:	M1 / E2
Temperature range	:	-25 °C / +55 °C

Further properties are described in the annexes:
– Description T10682 revision 5;
– Documentation folder T10682-5.

Valid until 28 May 2025

Remarks This revision replaces the earlier versions, excluding its documentation folder.

Issuing Authority

NMI Certin B.V., Notified Body number 0122
4 April 2019


C. Oosterman
Head Certification Board