

Issued by	NMi Certin B.V., designated and notified by the Netherlands to perform tasks with respect to conformity assessment procedures mentioned in article 17 of Directive 2014/32/EU, after having established that the Measuring instrument meets the applicable requirements of Directive 2014/32/EU, to:
Manufacturer	ITRON 1 Avenue des temps modernes ZI du Bernais B.P. 23 86361 Chasseneuil du Poitou France
Measuring instrument	A static Active Electrical Energy Meter Type : ITE411L3B Manufacturer's mark or name : ITRON Reference voltage : 230 V Reference current : 10 A Destined for the measurement of : electrical energy, in a - single-phase two-wire network Accuracy class : B Environment classes : M1 / E2 Temperature range : -25 °C / +55 °C Further properties are described in the annexes: - Description T10716 revision 7; - Documentation folder T10716-6.
Valid until	14 July 2026
Remarks	This revision replaces the earlier versions, excluding its documentation folder.

Issuing Authority

NMi Certin B.V., Notified Body number 0122
1 November 2018



C. Oosterman
Head Certification Board

NMi Certin B.V.
Hugo de Grootplein 1
3314 EG Dordrecht
The Netherlands
T +31 78 6332332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the manufacturer shall indemnify third-party liability.

The designation of NMi Certin B.V. as Notified Body can be verified at
<http://ec.europa.eu/growth/tools-databases/nando/>

Reproduction of the complete document only is permitted.