

Issued by	NMi Certin B.V., designated and notified by the Netherlands to perform tasks with respect to conformity modules mentioned in article 17 of Directive 2014/32/EU, after having established that the Measuring instrument meets the applicable requirements of Directive 2014/32/EU, to:
Manufacturer	Elster Solutions SAS 1 allée de Bruxelles Estrées Deniecourt (80200) France
Authorised representative	Elster Solutions GmbH Steinern Str. 19-21 55252 Mainz-Kastel Hessen Germany
Measuring instrument	A static Active Electrical Energy Meter Type : AS330L-G3 Manufacturer's mark or name : Elster Reference voltage : 230 V Reference current : 10 A Destined for the measurement of : electrical energy, in a - single-phase two-wire network Accuracy class : B Environment classes : M1 / E2 Temperature range : -40 °C / +70 °C Further properties are described in the annexes: - Description T10756 revision 1; - Documentation folder T10756-2.
Valid until	10 February 2026
Remark	This revision replaces the previous versions, including its documentation folder

Issuing Authority **NMi Certin B.V., Notified Body number 0122**
23 February 2017



C. Oosterman
Head Certification Board