

Issued by	NMi Certin B.V., designated and notified by the Netherlands to perform tasks with respect to conformity modules mentioned in article 17 of Directive 2014/32/EU, after having established that the Measuring instrument meets the applicable requirements of Directive 2014/32/EU, to:
Manufacturer	ELGAS, s.r.o. Semtínská 211 Pardubice Czech Republic
Measuring instrument	A Rotary Displacement Gas Meter Type : EMR EMR-Dual Manufacturer's mark or name : ELGAS, s.r.o. Destined for the measurement of : Gas volume Accuracy class : Class 1,0 or 1,5 (see the tables under 1.2.1 of the description) Environment classes : M1 / E2 Temperature range : -40 °C / +70 °C Further properties are described in the annexes: – Description T10774 revision 0; – Documentation folder T10774-1.
Valid until	23 June 2027

Issuing Authority **NMi Certin B.V., Notified Body number 0122**
23 June 2017



C. Oosterman
Head Certification Board

NMi Certin B.V.
Hugo de Grootplein 1
3314 EG Dordrecht
The Netherlands
T +31 78 6332332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the manufacturer shall indemnify third-party liability.

The designation of NMi Certin B.V. as Notified Body can be verified at <http://ec.europa.eu/growth/tools-databases/nando/>

Reproduction of the complete document only is permitted.