

Issued by NMI Certin B.V.,  
designated and notified by the Netherlands to perform tasks with respect to  
conformity modules mentioned in article 17 of Directive 2014/32/EU, after  
having established that the Measuring instrument meets the applicable  
requirements of Directive 2014/32/EU, to:

Manufacturer Networked Energy Services  
5215 Hellyer Avenue, Suite 150  
San Jose, CA 95138  
United States of America

Measuring instrument A static **Active Electrical Energy Meter**  
Type : Direct Connected Meter (DCM)  
83334-3xxxxx, 84334-3xxxxx, 85334-3xxxxx  
Transformer Connected Meter (TCM)  
83534-3xxxxx, 84534-3xxxxx, 85534-3xxxxx  
83524-3xxxxx, 84524-3xxxxx, 85524-3xxxxx

Manufacturer's mark or name : Networked Energy Services

Reference voltage : 3x110/190 V ... 3x230/400 V  
2x110 V ... 2x230 V  
110 V ... 230 V

Reference current : 5 A (DCM/TCM), 10 A and 20 A (DCM)

Destined for the measurement of : electrical energy, in a  
- three-phase four-wire network  
- two-phase three-wire network  
- single-phase two-wire network

Accuracy class : A, B, C

Environment classes : M1 / E2

Temperature range : -40 °C / +70 °C

Further properties are described in the annexes:

- Description T10916 revision 18;
- Documentation folder T10916-11.

Valid until 4 April 2027

Remark This revision replaces the earlier versions, except its documentation folder.

Issuing Authority **NMI Certin B.V., Notified Body number 0122**  
20 September 2022

Certification Board