

Issued by NMI Certin B.V.,
designated and notified by the Netherlands to perform tasks with respect to conformity modules mentioned in article 17 of Directive 2014/32/EU, after having established that the Measuring instrument meets the applicable requirements of Directive 2014/32/EU, to:

Manufacturer Shenzhen Kaifa Technology (Chengdu) Co., Ltd.
No. 99 Tianquan Rd.,
Hi-Tech Development Zone
Chengdu 611731
P.R. China

Measuring instrument A static **Active Electrical Energy Meter**

| | | |
|---------------------------------|---|--|
| Type | : | MA105H2CNL or MA105 |
| Manufacturer's mark or name | : | KAIFA |
| Reference voltage | : | 230 V |
| Reference current | : | 5 A |
| Destined for the measurement of | : | electrical energy, in a - single-phase two-wire network |
| Accuracy class | : | B or A |
| Environment classes | : | M1 / E2 |
| Temperature range | : | -25 °C / +70 °C |

Further properties are described in the annexes:
– Description T11301 revision 3;
– Documentation folder T11301-4.

Valid until 29 May 2028

Remark This revision replaces the earlier versions, including its documentation folder.

Issuing Authority **NMI Certin B.V., Notified Body number 0122**
11 November 2022

Certification Board