

Issued by NMI Certin B.V.,
designated and notified by the Netherlands to perform tasks with respect to
conformity assessment procedures mentioned in article 17 of Directive
2014/32/EU, after having established that the Measuring instrument meets
the applicable requirements of Directive 2014/32/EU, to:

Manufacturer SIAME
Zone Industrielle, Rue des Mathématiques
8030 Grombalia
Tunisia

Measuring instrument A static **Active Electrical Energy Meter**
Type : ME300SM-P, ME300SM-G, ME300SM
Manufacturer's mark or name : SIAME
Reference voltage : 3x230/400 V
Reference current : 5 A
Destined for the measurement : electrical energy, in a
of - three-phase four-wire network
Accuracy class : A or B
Environment classes : M1 / E2
Temperature range : -40 °C / +70 °C

Further properties are described in the annexes:
– Description T11765 revision 0;
– Documentation folder T11765-1.

Valid until 10 December 2029

Issuing Authority **NMI Certin B.V., Notified Body number 0122**
10 December 2019

Certification Board