

Issued by NMI Certin B.V.,  
designated and notified by the Netherlands to perform tasks with respect to conformity modules mentioned in article 17 of Directive 2014/32/EU, after having established that the Measuring instrument meets the applicable requirements of Directive 2014/32/EU, to:

Manufacturer Apator S.A.  
ul. Gdańska 4a, lok C4  
87-100 Toruń  
Poland

Measuring instrument **A static Active Electrical Energy Meter**

Type	: PICUS+ or PICUS
Manufacturer's mark or name	: Apator S.A.
Reference voltage	: 3x230/400 V or 230 V
Reference current	: 5 A
Destined for the measurement of	: electrical energy, in a - three-phase four-wire network - one-phase two-wire network
Accuracy class	: B or A
Environment classes	: M1 / E2
Temperature range	: -40 °C / +70 °C or -25 °C / +55 °C

Further properties are described in the annexes:  
– Description T11887 revision 3;  
– Documentation folder T11887-4.

Valid until 29 October 2030

Remark This revision replaces the earlier version, including for its documentation folder.

Issuing Authority **NMI Certin B.V., Notified Body number 0122**  
8 July 2022

Certification Board

**NMI Certin B.V.**  
Thijsseweg 11  
2629 JA Delft  
The Netherlands  
T +31 88 636 2332  
certin@nmi.nl  
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the manufacturer shall indemnify third-party liability.

The designation of NMI Certin B.V. as Notified Body can be verified at <http://ec.europa.eu/growth/tools-databases/nando/>

Reproduction of the complete document only is permitted.

This document is digitally signed and sealed. The digital signature can be verified in the blue ribbon at the top of the electronic version of this certificate.