

Issued by NMI Certin B.V.,
designated and notified by the Netherlands to perform tasks with respect to conformity assessment procedures mentioned in article 17 of Directive 2014/32/EU, after having established that the measuring instrument meets the applicable requirements of Directive 2014/32/EU, to:

Manufacturer Itron Indonesia- PT Mecoindo
EJIP Plot 6B-2
Lemah Abang, Bekasi 17550
Jawa Barat - Indonesia

Measuring instrument A static **Active Electrical Energy Meter**

Type : EM600 TYPE620
(part of SL7000 series)

Manufacturer's mark or name : Itron

Reference voltage : 3x57,7/100 V to 3x277/480 V
3x133/230 V to 3x277/480 V

Reference current : 1 A , 5 A (CT version)
5 A (DC version)

Destined for the measurement of : electrical energy, in a
- three-phase four-wire network
- three-phase three-wire network

Accuracy class : A or B or C

Environment classes : M1 / E2

Temperature range : -40 °C / +70 °C

Further properties are described in the annexes
– Description T12284 revision 4
– Documentation folder T12284-4

Valid until 16 March 2032

Initially issued 16 March 2022

Remarks This revision replaces the earlier versions, including its documentation folder.

Issuing Authority **NMI Certin B.V., Notified Body number 0122**
2 February 2024

Certification Board