

Issued by NMI Certin B.V.,  
designated and notified by the Netherlands to perform tasks with respect to  
conformity assessment procedures mentioned in article 17 of Directive  
2014/32/EU, after having established that the Measuring instrument meets  
the applicable requirements of Directive 2014/32/EU, to:

Manufacturer Landis+Gyr  
78 km National Road Athens  
Corinth GR-20100  
Greece

Authorised Representative Landis+Gyr Oy  
Paperitehtaankatu 9  
FI-40100 Jyväskylä  
Finland

Measuring instrument **A static single phase electricity meter**

Type : E360 AD1D...B2

Manufacturer's mark or name : Landis+Gyr

Reference voltage : 230 V

Reference current : 10 A

Destined for the measurement of : electrical energy, in a  
- single-phase two-wire network

Accuracy class : A or B

Environment classes : M1 / E2

Temperature range : -25 °C / +70 °C

Further properties are described in the annexes:

- Description T12516 revision 5;
- Documentation folder T12516-2.

Valid until 8 December 2032

Initially issued 8 December 2022

Remark This version replaces the previous version(s), excluding the documentation folder.

Issuing Authority **NMI Certin B.V., Notified Body number 0122**  
21 September 2023

Certification Board

**NMI Certin B.V.**  
Thijsseweg 11  
2629 JA Delft  
The Netherlands  
T +31 88 636 2332  
certin@nmi.nl  
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the manufacturer shall indemnify third-party liability.

The designation of NMI Certin B.V. as Notified Body can be verified at <http://ec.europa.eu/growth/tools-databases/nando/>

Reproduction of the complete document only is permitted.

This document is digitally signed and sealed. The digital signature can be verified in the blue ribbon at the top of the electronic version of this certificate.