

Issued by NMI Certin B.V.,
designated and notified by the Netherlands to perform tasks with respect to
conformity assessment procedures mentioned in article 17 of Directive
2014/32/EU, after having established that the Measuring instrument meets
the applicable requirements of Directive 2014/32/EU, to:

Manufacturer Meter&Control
Veljka Dugoševića 54
Belgrade
Serbia

Measuring instrument A static **Active Electrical Energy Meter**

Type	: F2-TD
Manufacturer's mark or name	: Meter&Control
Reference voltage	: 3x230/400 V
Reference current	: 5 A
Destined for the measurement of	: electrical energy, in a - three-phase four-wire network
Accuracy class	: A or B
Environment classes	: M1 / E2
Temperature range	: -40 °C / +70 °C

Further properties are described in the annexes:
– Description T12562 revision 3;
– Documentation folder T12562-2.

Valid until 14 April 2033

Initially issued 14 April 2023

Remark This revision replaces the earlier versions, excluding its
documentation folder.

Issuing Authority **NMI Certin B.V., Notified Body number 0122**
5 January 2024

Certification Board