

Issued by NMI Certin B.V.,
designated and notified by the Netherlands to perform tasks with respect to
conformity assessment procedures mentioned in article 17 of Directive
2014/32/EU, after having established that the measuring instrument meets
the applicable requirements of Directive 2014/32/EU, to:

Manufacturer Honeywell Automation India Ltd.
Plot No. 3
Gat No. 181
Village Fulgaon, Tal-Haveli
Pune 412216, Maharashtra
India

Measuring instrument An **electronic gas-volume conversion device (EVCD)**, intended to be
used for gas volume conversion as a sub-assembly (according to article 4 of
the MID) of a gas meter.

Type : EK205 & EK205-M
Manufacturer's mark or name : Elster
Conversion principle : T, PT or PTZ
Ambient temperature range : EK205: -25 °C ... +55 °C
EK205-M : -40 °C ... +55 °C
Designed for : non-condensing humidity
Environment classes : M2 / E2

The intended location for the instrument is open.

Further properties are described in the annexes:

- Description T12625 revision 0
- Documentation folder T12625-1.

Valid until 18 July 2033

Issuing Authority **NMI Certin B.V., Notified Body number 0122**
18 July 2023

Certification Board

NMI Certin B.V.
Thijsseweg 11
2629 JA Delft
The Netherlands
T +31 88 636 2332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision that
no liability is accepted and that the
manufacturer shall indemnify third-party
liability.

The designation of NMI Certin B.V. as Notified
Body can be verified at [http://
ec.europa.eu/growth/tools-databases/nando/](http://ec.europa.eu/growth/tools-databases/nando/)

Reproduction of the complete
document only is permitted.

This document is digitally signed
and sealed. The digital signature
can be verified in the blue ribbon at
the top of the electronic version of
this certificate.