

Issued by NMI Certin B.V.,
designated and notified by the Netherlands to perform tasks with respect to conformity modules mentioned in article 17 of Directive 2014/32/EU, after having established that the Measuring instrument meets the applicable requirements of Directive 2014/32/EU, to:

Manufacturer Landis+Gyr
30 Avenue du Président Auriol
BP 3150
03115 Montluçon Cedex
France

Measuring instrument A static **Active Electrical Energy Meter**

Type	: L19C1 - ZMG416
Reference voltage	: 3x230/400 V
Reference current	: 5 A
Destined for the measurement of	: electrical energy, in a - three-phase four-wire network
Accuracy class	: C
Environment classes	: M1 / E2
Temperature range	: -25 °C / +55 °C

Further properties are described in the annexes:
- Description T10213 revision 10;
- Documentation folder T10213-8.

Initially issued: 16 July 2009

Valid until 16 July 2029

Remark This revision replaces the earlier versions, including its documentation folder.

Issuing Authority **NMI Certin B.V., Notified Body number 0122**
3 May 2024

Certification Board

NMI Certin B.V.
Thijsseweg 11
2629 JA Delft
The Netherlands
T +31 88 636 2332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the manufacturer shall indemnify third-party liability.

The designation of NMI Certin B.V. as Notified Body can be verified at <http://ec.europa.eu/growth/tools-databases/nando/>

This document is digitally signed and sealed. The digital signature can be verified in the blue ribbon on top of the electronic version of this certificate.

Reproduction of the complete document only is permitted.