

Issued by NMI Certin B.V.,
designated and notified by the Netherlands to perform tasks with respect to conformity assessment procedures mentioned in article 17 of Directive 2014/32/EU, after having established that the measuring instrument meets the applicable requirements of Directive 2014/32/EU, to:

Manufacturer EDMIT LIMITED
47 Yishun Industrial Park A
Singapore 768724

Measuring instrument A static **Active Electrical Energy Meter**

Type : Mk6E
Manufacturer's mark or name : EDMIT Limited
Reference voltage : 3x57/100V – 3x240/415V
3x100 V – 3x290 V
Reference current : 1 A or 5 A
Destined for the measurement of : electrical energy, in a
- three-phase four-wire network
- two-phase three-wire network
- single-phase two-wire network

Accuracy class : A or B or C
Environment classes : M1 / E2
Temperature range : -40 °C / +70 °C

Further properties are described in the annexes:
– Description T10679 revision 5;
– Documentation folder T10679-3.

Valid until 13 March 2025

Initially issued 13 March 2015

Remark This revision replaces the earlier versions, except for its documentation folder.

Issuing Authority **NMI Certin B.V., Notified Body number 0122**
4 July 2024

Certification Board