

Issued by NMI Certin B.V.,
designated and notified by the Netherlands to perform tasks with respect to conformity assessment procedures mentioned in article 17 of Directive 2014/32/EU, after having established that the measuring instrument meets the applicable requirements of Directive 2014/32/EU, to:

Manufacturer Itron Indonesia - PT Mecoindo
EJIP Plot 6B-2, Lemah Abang
17550 Bekasi, Jawa Barat
Indonesia

Measuring instrument A static **Active Electrical Energy Meter**

Type : EM512 TYPE 700 JV1
Manufacturer's mark or name : Itron
Reference voltage : 220, 230 or 240 V
Reference current : 5 A
Destined for the measurement of : electrical energy, in a
- single-phase two-wire network
Accuracy class : B
Environment classes : M1 / E2
Temperature range : -25 °C / +55 °C

Further properties are described in the annexes:
- Description T10940 revision 3;
- Documentation folder T10940-3.

Valid until 20 February 2027

Initially issued 20 February 2017

Remark This revision replaces the earlier versions, including its documentation folder.

Issuing Authority **NMI Certin B.V., Notified Body number 0122**
7 November 2024

Certification Board