

Issued by NMI Certin B.V.,
designated and notified by the Netherlands to perform tasks with respect to conformity assessment procedures mentioned in article 17 of Directive 2014/32/EU, after having established that the measuring instrument meets the applicable requirements of Directive 2014/32/EU, to:

Manufacturer Landis + Gyr A.E.
78th km National Road Athens-Corinth
GR-201 00, Corinth
Greece

Measuring instrument A static **Active Electrical Energy Meter**

Type : E450 G3 IDIS (ZCXi1x0xQ S4)
E450 G3 IDIS (ZBXi1x0xQ S4)

Manufacturer's mark or name : Landis + Gyr

Reference voltage : 230 V

Reference current : 5 A or 10 A

Destined for the measurement of : electrical energy, in a
- single-phase two-wire network

Accuracy class : B or A

Environment classes : M2 / E2

Temperature range : -40 °C / +70 °C

Further properties are described in the annexes:

- Description T11169 revision 14;
- Documentation folder T11169-4.

Valid until 1 September 2027

Initially issued 1 September 2017

Remarks This revision replaces the earlier versions, including its documentation folder.

Issuing Authority **NMI Certin B.V., Notified Body number 0122**
10 June 2025

Certification Board