

Issued by NMI Certin B.V.,  
 designated and notified by the Netherlands to perform tasks with respect to  
 conformity assessment procedures mentioned in article 17 of Directive  
 2014/32/EU, after having established that the measuring instrument meets  
 the applicable requirements of Directive 2014/32/EU, to:

Manufacturer Landis+Gyr Olympus AG  
 Alte Steinhäuserstr. 18  
 CH-6330 Cham  
 Switzerland

Authorised representative Landis+Gyr Oy  
 Paperitehtaankatu 9  
 FI-40100 Jyväskylä  
 Finland

Measuring instrument A static **Active Electrical Energy Meter**

Type : E450 ZMXi3x0xQ... S5

Manufacturer's mark or name : Landis+Gyr

Reference voltage : 3x230/400V

Reference current : 5 A or 10 A

Destined for the measurement of : electrical energy, in a  
 - three-phase four-wire network  
 - two-phase three-wire network  
 - one-phase two-wire network

Accuracy class : A or B

Environment classes : M2 / E2

Temperature range : -40 °C / +70 °C

Further properties are described in the annexes:

- Description T11968 revision 14;
- Documentation folder T11968-9.

Valid until 27 October 2030

Initially issued 27 October 2020

Remark This revision replaces the earlier versions, including for its documentation folder.

Issuing Authority **NMI Certin B.V., Notified Body number 0122**  
 30 April 2026

Certification Board