

Issued by

NMi Certin B.V.,
designated and notified by the Netherlands to perform tasks with respect to conformity assessment procedures mentioned in article 17 of Directive 2014/32/EU, after having established that the Measuring instrument meets the applicable requirements of Directive 2014/32/EU, to:

Manufacturer

Landis+Gyr
78 km National Road Athens
Corinth GR-20100
Greece

Authorized representative

Landis+Gyr Oy
Paperitehtaankatu 9
FI-40100 Jyväskylä
Finland

Measuring instrument

A static poly phase electricity meter

Type : E360-AM3D...S2
E360-AM3D...s2

Manufacturer's mark or name : Landis+Gyr

Reference voltage : 230V; 2-3x230/400V

Reference current : 5 A or 10 A

Destined for the measurement of : electrical energy, in a
- three-phase four-wire network
- two-phase three-wire network
- single-phase two-wire network

Accuracy class : A or B

Environment classes : M1 / E2

Temperature range : -40 °C / +70 °C

Further properties are described in the annexes:

- Description T12072 revision 17;
- Documentation folder T12072-8.

Valid until

29 June 2031

Initially issued

29 June 2021

Remark

This revision replaces the earlier versions, including its documentation folder

Issuing Authority

NMi Certin B.V., Notified Body number 0122
11 June 2024

Certification Board