

Issued by NMI Certin B.V.,  
designated and notified by the Netherlands to perform tasks with respect to  
conformity assessment procedures mentioned in article 17 of Directive  
2014/32/EU, after having established that the Measuring instrument meets  
the applicable requirements of Directive 2014/32/EU, to:

Manufacturer Landis+Gyr  
78 km National Road Athens  
Corinth GR-20100  
Greece

Authorised representative Landis+Gyr Oy  
Paperitehtaankatu 9  
FI-40100 Jyväskylä  
Finland

Measuring instrument A static **Single Phase Active Electrical Energy Meter**  
Type : E360-AD2D...S2  
Manufacturer's mark or name : Landis+Gyr  
Reference voltage : 230 V  
Reference current : 5 A or 10 A  
Destined for the measurement of : electrical energy, in a  
- single-phase two-wire network  
Accuracy class : A or B  
Environment classes : M1 / E2  
Temperature range : -40 °C / +70 °C

Further properties are described in the annexes:  
– Description T12420 revision 8;  
– Documentation folder T12420-5.

Initially issued 4 January 2023

Valid until 4 January 2033

Remark This revision replaces the previous versions, except for its documentation folder.

Issuing Authority **NMI Certin B.V., Notified Body number 0122**  
11 April 2024

## Certification Board