

Issued by NMI Certin B.V.,
designated and notified by the Netherlands to perform tasks with respect to
conformity assessment procedures mentioned in article 17 of Directive
2014/32/EU, after having established that the Measuring instrument meets
the applicable requirements of Directive 2014/32/EU, to:

Manufacturer Landis+Gyr
78 km National Road Athens
Corinth GR-20100
Greece

Authorised Representative Landis+Gyr Oy
Paperitehtaankatu 9
FI-40100 Jyväskylä
Finland

Measuring instrument **A static single phase electricity meter**

Type : E360 AD1D...B2

Manufacturer's mark or name : Landis+Gyr

Reference voltage : 230 V

Reference current : 10 A

Destined for the measurement of : electrical energy, in a
- single-phase two-wire network

Accuracy class : A or B

Environment classes : M1 / E2

Temperature range : -25 °C / +70 °C

Further properties are described in the annexes:

- Description T12516 revision 7;
- Documentation folder T12516-3.

Valid until 8 December 2032

Initially issued 8 December 2022

Remark This version replaces the previous version(s), including the documentation folder.

Issuing Authority **NMI Certin B.V., Notified Body number 0122**
7 March 2025

Certification Board