

Issued by NMI Certin B.V.,  
designated and notified by the Netherlands to perform tasks with respect to conformity assessment procedures mentioned in article 17 of Directive 2014/32/EU, after having established that the measuring instrument meets the applicable requirements of Directive 2014/32/EU, to:

Manufacturer JHB Technology BV.  
Burgemeester Burgerslaan 40  
5245 NH Rosmalen  
The Netherlands

Measuring instrument A static **Active Electrical Energy Meter**

Type : IQCT04000br1  
Manufacturer's mark or name : JHB Technology BV.  
Reference voltage : 3x230/400V V  
Reference current : 1 A  
Destined for the measurement of : electrical energy, in a  
- three-phase four-wire network  
Accuracy class : C  
Environment classes : M1 / E2  
Temperature range : -40 °C / +70 °C

Further properties are described in the annexes:  
- Description T12755 revision 0;  
- Documentation folder T12755-1.

Valid until 28 March 2034

Initially issued 28 March 2024

Issuing Authority **NMI Certin B.V., Notified Body number 0122**  
28 March 2024

Certification Board