

Issued by NMI Certin B.V.,
designated and notified by the Netherlands to perform tasks with respect to conformity assessment procedures mentioned in article 17 of Directive 2014/32/EU, after having established that the measuring instrument meets the applicable requirements of Directive 2014/32/EU, to:

Manufacturer Eastron Europe Limited
Suite 20 Cornwallis House, Howard Chase
Basildon SS14 3BB
United Kingdom

Measuring instrument A static **Active Electrical Energy Meter**

Type : SDM54-DM

Manufacturer's mark or name : Eastron

Reference voltage : 3x230/400 V

Reference current : 5 A

Destined for the measurement of : electrical energy, in a
- single-phase two-wire network
- single-phase three-wire network
- three-phase four-wire network

Accuracy class : A or B

Environment classes : M1 / E2

Temperature range : -40 °C / +70 °C

Further properties are described in the annexes:

- Description T12775 revision 0;
- Documentation folder T12775-1.

Valid until 18 July 2034

Initially issued 18 July 2024

Issuing Authority **NMI Certin B.V., Notified Body number 0122**
18 July 2024

Certification Board