

EU-type examination certificate

Number **T12796** revision 2
Project number 3760889
Page 1 of 1

Issued by

NMi Certin B.V.,
designated and notified by the Netherlands to perform tasks with respect
to conformity assessment procedures mentioned in article 17 of Directive
2014/32/EU, after having established that the measuring instrument meets
the applicable requirements of Directive 2014/32/EU, to:

Manufacturer

Zhejiang Eastron Electronic Co., Ltd.
No.52, Dongjin Road, Nanhu
Jiaxing, Zhejiang
China

Measuring instrument

A **static DC energy meter** intended to be integrated in a EV charging
system

Type : DCM6-650 and DCM6-200

Manufacturer's mark or name : Eastron

Reference voltage : 150 VDC ... 1000 VDC

Reference current : 130 A (DCM6-650)
40 A (DCM6-200)

Destined for the measurement of : electrical energy, in a
- single-phase two-wire network

Accuracy class : A or B

Environment classes : M1 / E2

Temperature range : -40 °C / +70 °C

Further properties are described in the annexes:

- Description T12796 revision 2;
- Documentation folder T12796-2.

Valid until

28 June 2034

Initially issued

28 June 2024

Remark

This revision replaces the previous versions, including the docfolder.

Issuing Authority

NMi Certin B.V., Notified Body number 0122
17 March 2025

Certification Board