

Issued by NMI Certin B.V.,
 designated and notified by the Netherlands to perform tasks with respect to conformity assessment procedures mentioned in article 17 of Directive 2014/32/EU, after having established that the measuring instrument meets the applicable requirements of Directive 2014/32/EU, to:

Manufacturer Zhejiang Eastron Electronic Co., Ltd.
 No.52, Dongjin Road, Nanhu
 Jiaxing, Zhejiang
 China

Authorised representative Eastron Europe Limited
 Suite 20 Cornwallis House, Howard Chase
 Basildon SS14 3BB
 United Kingdom

Measuring instrument A static **Active Electrical Energy Meter**

Type : SDM230-CB

Manufacturer's mark or name : Eastron

Reference voltage : 230 V

Reference current : 10 A

Destined for the measurement of : electrical energy, in a
 - single-phase two-wire network

Accuracy class : A or B

Environment classes : M1 / E2

Temperature range : -40 °C / +70 °C

Further properties are described in the annexes:

- Description T12830 revision 1;
- Documentation folder T12830-1.

Valid until 18 July 2034

Initially issued 18 July 2024

Remark This revision replaces the previous version, except for its documentation folder.

Issuing Authority **NMI Certin B.V., Notified Body number 0122**
 17 June 2025

Certification Board