

EC TYPE EXAMINATION CERTIFICATE of a measuring instrument

Certificate Number NL – 1122 – MI005B – 2014 – 001

Issued to: Haar Nederland C.V., Van Polanenweg 1, 2921 LT Krimpen a/d IJssel.

Issued by: SGS Netherlands, P.O. Box 200, 3200 AE Spijkenisse, The Netherlands, being Notified Body 1122.

Issue date: 24 September 2014

In accordance with: Metrologiewet (Stb. 2006, 137) as dutch implementation of Council Directive 2004/22/EC on measuring instruments, annex MI-005, annex B: Type Examination.

In respect of: An interruptible measuring system for liquids other than water for deliveries of AdBlue from road trucks and tank containers, with the following main characteristics:

Manufacturer:	Haar Nederland
Type:	BS WG65 ELZ
Qmin - Qmax	70 – 500 l/min
Minimum measured quantity:	70 l
Accuracy class	0.5
Environmental classes:	E3, M3
For measurement of:	AdBlue
Temperature range env.:	-10 °C - +50 °C
Temperature range liquid:	-10 °C - +50 °C
Maximum pressure:	2.5 bar(g)

The necessary data (principal characteristics, approval conditions, securing, functioning, etc.) for identification purposes and conditions (when applicable) are set out in the descriptive annex to this certificate. All documentations are filed under reference number 919239.

Authorization: M. van der Flier, Technisch Manager Statutory Services,

For and on behalf of
SGS Nederland B.V.

